


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	BRE GLOBAL LIMITED Incorporating LPC Testing, BRE Testing Issue No:067 Issue date: 25 June 2010	
	Bucknalls Lane Garston Watford WD25 9XX	Contact: Mr J A Noonan Tel: +44 (0)1923 664334 Fax: +44 (0)1923 664994 E-Mail: noonanj@bre.co.uk Website: www.bre.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Bucknalls Lane Garston Watford WD25 9XX Local contact: Mr J A Noonan	Acoustics Construction Fire resistance & protection Reaction to fire Electrical EMC, IT, Radio & Telecoms Airborne organic compounds Wood & wood preservatives Doors and Windows	A
Address BRE GLOBAL LIMITED Incorporating LPC Testing, BRE Testing - Testing facility Dawsons Wharf Riverside Park Road Middlesbrough TS2 1UT Contact: Mr J A Noonan	Construction Fire resistance & protection Reaction to fire Doors	C

Site activities performed away from the locations listed above:

Location details	Activity	Location code
All locations suitable for the activities listed Contact: Mr J A Noonan	Acoustics Construction Environmental sampling Fire resistance & protection Environmental Sampling	B



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
APPLIED PROTECTION TO STRUCTURAL MEMBERS	Fire resistance	pr EN 13381-8:2008 pr EN 13381-4:2008 DD ENV 13381-4:2002 (Test Method only) DD ENV 13381-2: 2002 DD ENV 13382-3: 2002 Pr EN 133281-4:2009 DD ENV 13382-5: 2002 DD ENV 13382-6: 2002 DD ENV 13382-7: 2002 Pr EN 13381-8:2007 (dated 04/2009)	A
BEAMS	Fire resistance	BS 476:Part 8:1972 Section 5 BS 476:Part 21:1987 Section 5 FTSG Resolutions 47 & 70 BS EN 1365-3:2000 ISO 834:1975	A
BUILDING ELEMENTS - GENERAL	Fire resistance	BS 476:Part 8:1972 BS 476:Part 20:1987 BS EN 1363-1:1999 BS EN 1363-2:1999 AS 1530:Part 4:1997 AS 1530:Part 4:2008 Hydrocarbon DoE FRS 14-84:1990 ISO 834-1:1999 FTSG. Resolutions 47, 51, 53, 68, 70, 71, 77, 78, 79 ASTM E119:2008a UL 203	A
	Classification using data from fire resistance tests excluding ventilation services	BS EN 13501-2:2007+A1:2009	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
BUILDING ELEMENTS	<u>Acoustical Tests</u> Measurement of airborne sound insulation	BS EN ISO 140-3:1995 BS 2750:Part 3:1995	A
	Field measurement of airborne sound insulation	BS EN ISO 140-4:1998 ISO 140-4:1998	B
BUILDING ELEMENTS - small	<u>Acoustical Tests</u> Measurement of airborne sound insulation	BS EN 20140-10:1992 ISO 140-10:1991	A
BUILDING MATERIALS/PRODUCTS	<u>Fire Tests</u> Calorimeter measurements	Documented In-House Method No SP8/12/T:1999 and SP8/13/T:1999	A
	Reaction to fire	BS EN ISO 1182:2002 BS EN ISO 11925-2:2002 BS EN 13823:2002 prEN SBI:1999 BRE Fire Note 9:1999 BS 8414-1:2002 BS 8414-2:2005	A
	Non-combustibility	BS 476:Part 4:1970(1984) ISO 1182:1990 AS 1530:Part 1:1994	A
	Heat emission from buildings	BS 476:Part 11:1982(1988)	A
	Ignitability	BS 476:Part 12:1991 BS 476:Part 13:1987	A
	Fire propagation in materials and products	BS 476:Part 6:1989 + A1:2009	A
	Surface spread of flame	BS 476:Part 7:1997 FTSG Resolution 81	A
	Specific optical density of smoke from solids	BS 6401:1983 ASTM/E 662/97 ASTM/E 662/03 ASTM/E 662/06	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
BUILDING MATERIALS/PRODUCTS (cont'd)	Gross calorific value (Bomb calorimeter test)	BS EN ISO 1716:2002	A
	Single burn item	BS EN 13823:2002	A
	Calorimeter measurements	Documented In-house Method No SP8/12/T:1999 and SP8.13/T:1999	A
	Room test	BS 476-33: 1993, ISO 9705:1993 UBC 26-3(1997)	A
	Cone calorimeter	BS 476-15: 1993 ISO 5660-1:1993 ASTM E 1354: 1997 ISO5660-1 & 2: 2002 prCEN/TS 45545-2008	A
	Naval Engineering Standards: Smoke Index (Category 2)	Def Stan 711-Issue 2:2006 NES 711	A
	Toxicity Index (Category 2)	NES 713 Def Stan 713-Issue 2:2006	
	Oxygen Index (Category 2)	NES 714 ISO 4589-2:1999 (Inc Amd 1)	
	Temperature Index (Category 2)	NES 715 ISO 4589-3:1996	
	<u>European Harmonised Fire Test</u>		
Conditioning	BS EN ISO 13238:2001	A	
Gross Calorific value (Bomb calorimeter test)	BS EN ISO 1716:2002		
Non-combustibility	BS EN ISO 1182:2002		
Small Flame test	BS EN ISO 11925-2:2002		
Single burning item test	BS EN 13823:2002		
Room corner test	BS EN 14390:2007		
Reaction to fire classification	BS EN 13501-1:2002 BS EN 13501-1:2007		



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
BUILDING MATERIALS/PRODUCTS (cont'd)	Reaction to fire	BS 8414-1:2002 BS 8414-2:2005 FM Approval Standard 4880 Fire performance of external thermal insulation for walls of multi-storey buildings Annex A: Building Research Standard 135:2003	A&C
	Room test	UBC 26-3(1997)	C
	Railway Applications	CEN TS 45545-2:2006: final draft A BS 6853:1999 (inc Amd 1)	C
BUTTERFLY STOP VALVES	<u>Performance Tests</u>	Documented In-house Method LPS 1185:Issue 3:1993	A
CABLES and MATERIALS	Oxygen index at elevated temperatures	BS EN ISO 4589-3:1996 LUL 2 - 01001 - 002: A1: 2003	A
	Determination of burning behaviour by Oxygen Index	BS EN ISO 4589-2:1999 ISO 4589:1996 ASTM D 2863-00 ASTM D 2863-06A	A
	Smoke density	BS 6853:1999 Annex D BS EN ISO 5659-2:1999 prEN 3475-602:1999 BSS 7324 prEN 2824-6 ATS 1000.01 Issue 5 Clause 6.2 LUL 1-085:A1:2008	A
	Emissions, Toxicity and Density	prEN 3475-601:1999 LPS 1207:1994 LPS 1207:issue 2:2003 BSS 7239 BSS 7238 BMS 13-55D BS 6853:1999 Annex B1 & B2 LUL 1-085:A1:2008	A
	Railway Applications	CEN TS 45545-2:2006: final draft A	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
CEILING MEMBRANES	Fire resistance - horizontal	BS 476:Part 8:1972 Section 9 BS 476:Part 22:1987 Section 9 BS EN 1364-2:1999 FTSG Resolutions 70	A
COLUMNS	Resistance to Fire	BS 476:Part 8:1972 Section 4 BS 476:Part 21:1987 Section 6 FTSG Resolution 14 and 70 BS EN 1365-4:2000 ISO 834:1975	A
COVERINGS	Fire resistance	BS EN 14135: 2004	A
DOOR and SHUTTER ASSEMBLIES	Fire resistance	BS 476:Part 8:1972 Section 7 BS 476:Part 22:1987 Section 6, 7 & 8 FTSG Resolutions 51, 70, 71, 78 & 79 BS EN 1634-1:2008 ISO 834:1975 ISO 3008:1978 ISO/DIS 3008: 2003 BS EN 81-58: 2003	A
	Opening and closing test	Documented In-House Methods LPS 1056 Appendix B	A
	Intumescent seal contribution	BS ISO 12472: 2003 BS 476-23:1987 Method 6	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
DOOR ASSEMBLIES - external, single leaf	<u>Physical Tests</u> Abusive forces on handles Basic security test Closure against obstruction Cyclic operation test Operating forces Resistance to hard body impact Resistance to static torsion Resistance to soft and heavy body impact Resistance to vertical loads Slamming resistance Weathertightness	PAS 23-1:1999	A
DOOR ASSEMBLIES - external, single and double leaf	Hard body impact Infill medium removal - manual Infill medium removal - mechanical Manual check test Manipulation test Mechanical loading test Soft body impact Manual cutting test Security hardware and cylinder test (Parts 1 & 2)	PAS 24-1:2007+A1:2009	A
DOOR AND WINDOW	Air Permeability	BS EN 1026:2000	A
	Watertightness	BS EN 1027:2000	A
	Resistance to wind loads	BS EN 12211:2000	A
	Weathertightness	BS 6375:Part 1:2004	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ELECTRICAL CABLES	<u>Fire Tests</u>		
	Railway Applications	CEN TS 45545-2:2006	A
	Smoke emission	BS 7211:1998 Clause 15 BS 7211:1994, Appendix D IEC 61034-1:1997 IEC 61034-1:2005 BS EN 61034-1:2005 BS EN 50268-2:2000 LUL 2 - 01001 - 002: A1: 2003 IEC 61034-2:1997 IEC 61034-2:2005 BS EN 61034-2:2005 BS 6724:1990 BS 6853:1987, Appendix B BS 6853:1999, Appendix D BS 7622-1:1993 BS EN 7622-2:1993 BS EN 50268-1:2000 LUL E4156: A4, Clause 10.4: 2000 BS EN 50268-1:2000 BS EN 50306-1:2002 Clause 8.3 BS EN 50306-2:2002 Clause 4:3 BS EN 50306-3:2002 Clause 4:3 BS EN 50306-4:2002 Clause 4:3 AS/NZS 1660.5.2:2005 LUL 1-085:A1:2008	A
<u>Fire Tests</u>			
	Acid & corrosive gas emission	BS EN 50267-2-1:1999 IEC 60754-1:1994 BS EN 50267-2-2:1999 BS EN 50267-2-3:1999 IEC 60754-2:1997 AS/NZS 1660 5.3:2005 AS/NZS 1660.5.4:2005 BS 6425:Part 1:1990	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ELECTRICAL CABLES (cont'd)	<u>Fire Tests</u>		
	Fire performance (vertical tests)	IEC 60332-3-10:2008 BS 4066:Part 3:1994 IEC 332-3: 1992 LUL 2 - 01001 - 002: A1: 2003 IEC 60332-3-21:2000 BS EN 50266-2-1:2001 IEC 60332-3-22:2008 BS EN 50266-2-2:2001 IEC 60332-3-23:2008 BS EN 50266-2-3:2001 IEC 60332-3-24:2008 BS EN 50266-2-4:2001 IEC 60332-3-25:2008 BS EN 50266-2-5:2001 IEC 60332:Part 3:1992 LUL E4156: A4, Clause 10.3:2000 BS EN 50305:2000 Clause 9:1 BS EN 50306-1:2002 Clause 8:2 BS EN 50306-2:2002 Clauses 4:1a & 4:1b BS EN 50306-3:2002 Clauses 4:1, 4:1b & 4:1c BS EN 50306-4:2002 Clauses 4:1a,4.1b,4.1c & 4:1d AS/NZS 1660.5.1:2005 AS/NZS 1660.5.6:2005 LUL 1-085:A1:2008	A
	Fire-resisting characteristics of electric cables	IEC 331:1970	A
Vertical flame propagation on a single wire cable	BS 4066-2:1980 BS 4066:Part 1:1980 IEC 60332-1-2:2004 BS EN 50265-2-1:1999 IEC 60332-1-3:2004 BS EN 50265-2-3:1999 IEC 60332-2-2:2004 BS EN 50265-2-2:1999 BS 6387:1994 Clause 10	A	



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ELECTRICAL CABLES (cont'd)	<u>Performance Tests</u>		
	Bending	BS 6387:1994 Clause 8	A
	Impact	Clause 9	
	Resistance to fire alone	Clause 11.1	
	Resistance to fire with water	Clause 11.2	
	Resistance to fire with mechanical shock	Clause 11.3	
	Resistance to fire alone	LUL E4156: A4, Clause 10.6, Category C: 2000 LUL 2-01001-005:2006	A
	Resistance to fire with water	LUL E4156: A4, Clause 10.6, Category W: 2000 LUL 2-01001-005:2006	A
	Resistance to fire with mechanical shock	LUL E4156: A4, Clause 10.6, Category Z: 2000 LUL 2-01001-005:2006	A
	<u>Performance Tests</u>		
Bending	BS 7629-1:2008 Clause 14.2	A	
Impact	Clause 14.3		
Resistance to fire with mechanical shock and with water	Clause 14.4		
Resistance to fire alone	BS 7629-1997 Clause 16.2	A	
Bending	Clause 16.3		
Impact	Clause 16.4		
Resistance to fire with water	Clause 16.5		
Resistance to fire with mechanical shock			
Resistance to fire	BS 7846:1996 Clause 19.6	A	
Resistance to fire - Category F1	BS 7846:2000 Clause 17.4.2	A	
Category F2	Clause 17.4.3		
Category F3	Clause 17.4.4		



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ELECTRICAL CABLES (cont'd)	Circuit Integrity	BS EN 50200:2000 BS EN 50200:2006 BS 8434-1:2003 BS 8434-2:2003 BS 5839-1:2008 [clause 26.2 (d) & (e)] BS 7346-6:2005 BS 8491:2008 AS/NZS 1660.5.5.:2005	A
	<u>Performance Tests</u> Resistance to fire	LUL 2-01001-005:2006 Clause 3.2.1.6.1 IEC60331-21:1999 IEC60331-23:1999 IEC60331-25:1999 IEC60331-31:2002	A
	Specification for mineral insulated cables	BS 6207-1:1995 Clause 7.2.5 Clause 7.3.1 Clause 7.3.2 Clause 8.3 BS EN 60702-1:2002 Clause 12.4 Clause 12.5 Clause 12.6 Clause 13.8	A
	Visual Examination Inspection	BS 7629-1:2008 (clauses 7,8.1,8.2 & 11) BS 7629-1:1997 (amendment 1 & 2) Clause 7, 8:1, 11 BS 7846-2000 Clause 7, 8.1, 12 BS 60702:2002 Clause 9	A
	Conductor resistance	BS 6360:1991 BS 6360:2005 BS EN 60228:2005 BS 7629-1:2008 (clause 13.2) BS 7629-1:1997 (Clause 15.3) BS 7629-2:1997 (Clause 15.3) BS 7846:2000 (Clause 16.2)	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
EXTERNAL CLADDING SYSTEMS	<u>Fire Tests</u> Fire performance & classification	Fire performance of external thermal insulation for walls of multi-storey buildings Annex A: Building Research Standard 135:2003 BRE Fire Note 9:1999 BS 8414-1:2002 BS 8414-2:2005 BR 135:annex B:DG501	A
FACADES	<u>Acoustical Tests</u> Field measurement of airborne sound insulation	BS EN ISO 140-5:1998	B
FIRE DETECTION & ELECTRONICS TESTING	Fire detection and fire alarm systems Heat sensitive detectors - Point detectors containing static element Point-type smoke detectors using scattered light, transmitted light or ionisation Control and indicating equipment Fire alarm devices - Sounders	BS 5445:Part 5:1977 (AMD 5762) EN 54:Part 5:1976 (+ A1:1998) BS EN 54-5:2001 (incorporating Amd No 1) EN 54-5:2000 ISO 7240-5:2003 BS 5445:Part 7:1984 (AMD 5763) EN 54:Part 7:1982 (+ A1:1998) BS EN 54-7:2001 incorporating Amd No 1 + A2 (amd 16539, 29 Sept 2006) EN 54-7:2000 ISO 7240-7:2003 BS EN 54-2:1998 + A1 (amd 16781, 30 Nov 2006 & A2 (amd 16967, 31 Jan 2007) EN 54-2:1997 ISO 7240-2:2003 BS EN 54-3:2001 (incorporating Amd No 1 + A 2 (amd 16462, 30 June 2006)) EN 54-3:2001	A A A A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (Cont'd)	Components of voice alarm systems - Loudspeakers	BS EN 54-24:2008 EN 54-24: 2008	A
	Power supply equipment	BS EN 54-4:1998 (incorporating Amd No 1 & 2) EN 54-4:1997 BS ISO 7240-4:2003	A
	Manual call points	BS 5839:Part 2:1983 (AMD 4685 & 5085) EN 54-11:2001	A
	Aspirating smoke detectors - Point Detectors	EN 54-20:2006	A
	Requirements and test methods for multisensor detectors, which respond to smoke and heat, and smoke detectors with more than one smoke sensor	CEA 4021:1999-06 (en) CEA 4021 July 2003 (en) BS ISO 7240-15:2004	A
	Short circuit isolators and Input/output devices	CEA GEI 1-052:1997 CEA GEI 1-084:2000 BS EN 54-17:2005 BS EN 54-18:2005	A
	Optical smoke detectors	BS EN 54-12:2002 EN 54-12:2002	A
	Fire alarm systems, flame detectors, point detectors	BS EN 54-10:2002	A
	Smoke Alarms		
	Point-type smoke alarms	BS 5446-1:2000	A
	Smoke alarm devices	BS EN 14604:2005	
Gas Detectors			
Carbon monoxide detectors	BS 7860:1996 (AMD 1 (10451)) BS EN 50291:2001	A	



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
	Shock - half sine, sawtooth, trapezoidal (at ambient temperatures) Maximum force: 106kN Maximum acceleration: 2207ms ⁻² (225g _n) Maximum velocity: 3.8ms ⁻¹ Maximum displacement: 50.8mm (p-p)	BS 2011:Ea:1987 BS EN 60068-2-27:1993:Ea IEC 68-2-27:1987 (1993)	A
	Free Fall Drop Max ht: 2 m Max drop area: 0.5 m x 0.5 m	IEC 60068-2-32:1975 + A1:1982 + A2:1990 Proc 1	A
	Impact (Spring Hammer)	IEC 60068-2-75:1997 Test Ehb	A
Fire detection and fire alarm systems	Bump - half sine (at ambient temperatures) Maximum force: 106kN Maximum acceleration: 2207ms ⁻² (225g _n) Maximum velocity: 3.8ms ⁻¹ Maximum displacement: 50.8mm (p-p)	BS 2011:Eb:1987 BS EN 60068-2-29:1993 (AMD 7827) IEC 68-2-29:1987 (1993)	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (cont'd)	<p><u>EMC</u></p> <p>1.1 Conducted emissions, 9kHz to 30MHz</p> <p>Performed in 4m x 4m x 2.4m unlined screened room</p> <p>Excluding discontinuous interference</p>	<p>EN 55011:1991 BS EN 55011:1991 CISPR 11:1990 (AMD 2:1996) EN 55011:1998 (AMD 1) BS EN 55011:1998 (AMD 1) CISPR 11:1997 (AMD 1:1999)</p> <p>EN 55014-1:1993 (AMD 1 & 2) BS EN 55014-1:1997 BS EN 55014-1:2001 EN 55014-1:2000 CISPR 14-1:2000</p> <p>EN 55022:1994 (AMD 1 & 2) BS EN 55022:1995 (AMD 1 & 2) EN 55022:1998 (AMD 1) BS EN 55022:1998 (AMD 1) CISPR 22:1997 (AMD 1)</p> <p>AZ/NZS 3548:1995 (AMD 1 & 2)</p>	<p>A</p> <p>A</p> <p>A</p> <p>A</p>
	<p>1.2 Radiated emissions 150kHz to 1000MHz</p> <p>Performed in 9m x 6m x 6.5m semi-anechoic chamber at 3m distance</p>	<p>EN 55011:1991 BS EN 55011:1991 CISPR 11:1990 (AMD 2:1996) EN 55011:1998 (AMD 1) BS EN 55011:1998 (AMD 1) CISPR 11:1997 (AMD 1:1999)</p>	<p>A</p>



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (cont'd)	<p><u>EMC</u></p> <p>Tests for automotive components performed in the semi-anechoic chamber</p>	<p>EN 55022:1994 (AMD 1 & 2) BS EN 55022:1995 (AMD 1 & 2) EN 55022:1998 (AMD 1) BS EN 55022:1998 (AMD 1) CISPR 22:1997 (AMD 1)</p> <p>AZ/NZS 3548:1995 (AMD 1 & 2)</p> <p>Directive 95/54/EC:1995 (Annex VII & VIII)</p> <p>UN Regulation No.10, Add.9/Rev 2:1997 (Annex 7 & 8)</p> <p>Directive 97/24/EC:1997 Chapter 8</p>	<p>A</p> <p>A</p> <p>A</p> <p>A</p>
	<p>1.3 Radiated immunity (GTEM)</p> <p>10kHz to 1GHz up to 100V/m</p> <p>1GHz to 2GHz up to 30V/m performed in 10m x 4m x 3m GTEM cell</p> <p>Closed loop or substitution method</p>	<p>BS 6667:Part 3:1985 IEC 801-3:1984</p> <p>ENV 50140:1994</p> <p>EN 61000-4-3:1996 (AMD 1 & 2) BS EN 61000-4-3:1997 (AMD 1 & 2) IEC 1000-4-3:1995 (AMD 1 & 2)</p> <p>ENV 50204:1995</p> <p>ISO 11452-3-2001</p> <p>Thatcham Performance test Specification Issue 2:1996</p>	<p>A</p> <p>A</p> <p>A</p> <p>A</p> <p>A</p> <p>A</p>



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (cont'd)	<u>EMC</u>		
	1.4 Radiated immunity (3m Anechoic Chamber)	BS 6667:Part 3:1985 IEC 801-3:1984	A
	1MHz to 2000MHz up to 30V/m	ENV 50140:1994	A
	1GHz to 2GHz up to 20V/m	ENV 50204:1995	A
		EN 61000-4-3:1996 (AMD 1 & 2) BS EN 61000-4-3:1997 (AMD 1 & 2) IEC 1000-4-3:1995 (AMD 1 & 2)	A
	Automotive: 200MHz to 1GHz up to 75 V/m	ISO 11452-2:1995	A
	Performed in 9m x 6m x 6.5m semi-anechoic chamber	Directive 95/54/EC:1995 (Annex IX) UN Regulation No. 10, Add.9/Rev 2:1997 (Annex 9)	A
	Closed loop or substitution method	Directive 97/24/EC:1997 (Chapter 8)	A
	1.5 Conducted RF Immunity 30kHz to 500MHz	ENV 50141:1994	A
	To Injected current 60mA	EN 61000-4-6:1996 (AMD 1 & 2) BS EN 61000-4-6:1996 (AMD 1 & 2) IEC 1000-4-6:1996 (AMD 1 & 2) ISO 11452-4:2001	A
	Closed loop or substitution method	Directive 95/54/EC:1995 (Annex IX)	A
	Performed in screened enclosures	UN Regulation No.10, Add.9/ Rev 2:1997 (Annex 9) Directive 97/42/EC:1997 (Chapter 8)	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (cont'd)	<u>EMC</u>		
	1.6 Electrostatic Discharge (ESD) Immunity	BS 6667:Part 2:1985 IEC 801-2:1984	A
	To ± 16.5kV air and to ± 9kV contact discharge	IEC 801-2:1991 EN 60801-2:1993 BS EN 60801-2:1993	A
		EN 61000-4-2:1995 (AMD 1 & 2) BS EN 61000-4-2:1995 (AMD 1 & 2) IEC 1000-4-2:1995 (AMD 1 & 2)	A
		ISO TR 10605:1994 (Component testing)	A
	1.7 Immunity to Fast Transient Bursts	IEC 801-4:1988	A
	0.25kV to 4.0kV positive and negative polarity	EN 61000-4-4:1995 (AMD 1) BS EN 61000-4-4:1995 (AMD 1) IEC 1000-4-4:1995 (AMD 1)	A
	0.25kV to 4.0kV, positive and negative polarity		
	AC Mains single phase to 16A		
	1.8 Immunity to High Energy Surges	IEC 801-5 (draft):1993 ENV 50142:1994	A
To ±8kV open circuit voltage with 1.2/50 µs pulse	BS EN 61000-4-5:1995 (AMD 1 & 2) EN 61000-4-5:1995 (AMD 1) IEC 1000-4-5:1995 (AMD 1)	A	
To ±4kA short circuit current with 8/20 µs pulse		A	
AC Mains single and three phase to 25A		A	



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (cont'd)	<u>EMC</u> 1.9 Mains Supply Voltage Variations and Dips Voltages to 600V, current to 16A	IEC 1000-4-11:1994 IEC 61000-4-11:1994	A A
		1.10 Generic Standards	EN 50081-1:1992 BS EN 50081-1:1992 EN 61000-6-3:2001 BS EN 61000-6-3:2001 EN 50081-2:1993 BS EN 50081-2:1994 EN 61000-6-4:2001 BS EN 61000-6-4:2001 EN 50082-1:1992 (Informative) BS EN 50082-1:1992 (Informative) EN 50082-1:1997 BS EN 50082-1:1998 EN 61000-6-1:2001 BS EN 61000-6-1:2001 EN 50082-2:1995 BS EN 50082-2:1995 EN 61000-6-2:2001 BS EN 61000-6-2:2001
	1.11 Product Specific Standards	EN 50130-4:1995 (AMD 1&2) BS EN 50130-4:1996 (AMD 1) EN 50270:1999 BS EN 50270:1999 EN 55011:1991 BS EN 55011:1991 CISPR 11:1990 (AMD 2:1996) EN 55011:1998 (AMD 1) BS EN 55011:1998 (AMD 1) CISPR 11:1997 (AMD 1:1999)	A A A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (cont'd)	EMC (cont'd) 1.11 Product Specific Standards (cont'd)	EN 55014-1:1993 (AMD 1 & 2) BS EN 55014-1:1997 BS EN 55014-1:2001 EN 55014-1:2000 CISPR 14-1:2000	A
		EN 55022:1994 (AMD 1 & 2) BS EN 55022:1995 (AMD 1 & 2) EN 55022:1998 (AMD 1) BS EN 55022:1998 (AMD 1) CISPR 22:1997 (AMD 1)	A
		AZ/NZS 3548:1995 (AMD 1 & 2)	A
		EN 55014-2:1997 BS EN 55014-2:1997 CISPR 14-2:1997	A
		EN 55024:1998 BS EN 55024:1998 CISPR 24:1997	A
		EN 60601-1-2:1993 (AMD 1) BS EN 60601-1-2:1993 (AMD 1)	A
		EN 61326:1997 (AMD 1 & 2) BS EN 61326:1998 (AMD 1 & 2)	A
		EN 50121-4:2000 BS EN 50121-4:2000	A
		DNV Certification Notes No. 2.4:1995	A
		Lloyds Register Test Specification No 1:2002	A
Germanischer Lloyds Regulations for Performance of Type Tests Part 1:1997	A		
Germanischer Lloyds Regulations for Performance of Type Tests Part 1:2003	A		



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE DETECTION & ELECTRONICS TESTING (cont'd)	EMC (cont'd) 1.11 Product Specific Standards (cont'd)	Thatcham Performance Test Specification Issue 2:1996	A
		Thatcham Performance Test Specification for Motorcycles Issue 1:1999	A
FIRE FIGHTING SYSTEMS (BLANKETS)	<u>Construction and Performance Tests</u>	BS EN 1869:1997	A
FIRE FIGHTING SYSTEMS (REEL HOSES)	<u>Construction and Performance Tests</u>	BS 3169:1986	A
FIRE FIGHTING SYSTEMS (GAS COMPONENTS) Performance Tests for Container Valves	Valve body strength Leakage Operation at minimum and maximum pressure Stress corrosion Salt mist corrosion	Documented In-house Method LPS 1024:Issue 2:1987(1992) Clause 5.2 Clause 5.4 Clause 5.5 Clause 5.8.1 Clause 5.8.2	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE FIGHTING SYSTEMS (GAS COMPONENTS) (Cont'd) Requirements and test methods for non-electrical automatic control and delay devices	General design Connection threads Signal processing and indication System delay Function at ambient temperature Internal pressure Strength Operational reliability Corrosion Stress corrosion Vibration Pneumatic energy supply using pilot containers Operating forces Functional reliability Safety Classification Test Facility Documentation Marking and data	BS EN 12094-2:2003 Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.5 Clause 4.6 Clause 4.7 Clause 4.8 Clause 4.9 Clause 4.10 Clause 4.11 Clause 4.12 Clause 4.13 Clause 4.14 Clause 4.15 Clause 4.16 Clause 4.17 Clause 6	A
	Requirements and test methods for manual triggering and stop devices	BS EN 12094-3:2003 Clause 4.1 Clause 4.2 Clause 4.3 Clause 6	



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>FIRE FIGHTING SYSTEMS (GAS COMPONENTS) (Cont'd)</p> <p>Requirements and test methods for container valve assemblies and their actuators</p>	<p>Compliance General design Connection threads Function and ambient temperature Resistance to internal pressure Strength Leakage Operational Reliability Corrosion Stress corrosion Vibration resistance Operating force Functional reliability Manual actuators Documentation Marking</p>	<p>BS EN 12094-4:2004</p> <p>Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.5 Clause 4.6 Clause 4.7 Clause 4.8 Clause 4.10 Clause 4.11 Clause 4.12 Clause 4.14 Clause 4.15 Clause 4.16 Clause 4.17 Clause 6</p>	A
<p>Requirements and Testing High and Low Pressure Selector Valves and their Actuators for all gases</p>	<p>General design Connection threads & flanges Function and ambient temperature Resistance to internal pressure & leakage Resistance to bursting Operational reliability Resistance to corrosion Resistance to stress corrosion Vibration resistance Operating force Documentation Marking</p>	<p>BS EN 12094-5:2006</p> <p>Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.5 Clause 4.6 Clause 4.8 Clause 4.9 Clause 4.10 Clause 4.11 Clause 4.12 Clause 6</p>	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE FIGHTING SYSTEMS (GAS COMPONENTS) (cont'd) Requirements and test methods for non-electrical disable devices Requirements and test methods for nozzles for CO2 systems Requirements and test methods for Connectors for CO2 Systems	General design Disabling valves Disable devices other than valves Operating forces Documentation Marking	BS EN 12094-6:2006 Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.5 Clause 6	A
	General design Connection threads Nozzle opening cross section Nozzle protection covers Distribution Resistance to pressure and heat Resistance to corrosion Resistance to stress corrosion Resistance to vibration Documentation Marking	BS EN 12094-7:2001 Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.6 Clause 4.7 Clause 4.9 Clause 4.10 Clause 4.11 Clause 4.12 Clause 6	A A
	General design Connection threads Resistance to leakage Resistance to bursting Resistance to cold Marking Documentation	BS EN 12094-8:2006 Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.7 Clause 4.9 Clause 4.10	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>FIRE FIGHTING SYSTEMS (GAS COMPONENTS) (cont'd)</p> <p>Requirements and test methods for special fire detectors</p>	<p>General</p> <p>Normal operating temperature</p> <p>Operating temperature</p> <p>Thermal response</p> <p>Pressure</p> <p>Force</p> <p>Function</p> <p>Heat Exposure</p> <p>Strength of release elements</p> <p>Strength of mounting support</p> <p>Thermal shock</p> <p>Resistance to low temperature</p> <p>Corrosion</p> <p>Stress corrosion</p> <p>Vibration resistance</p> <p>Documentation</p> <p>Marking</p>	<p>BS EN 12094-9:2003</p> <p>Clause 4.1</p> <p>Clause 4.2</p> <p>Clause 4.3</p> <p>Clause 4.4</p> <p>Clause 4.5</p> <p>Clause 4.6</p> <p>Clause 4.7</p> <p>Clause 4.8</p> <p>Clause 4.9</p> <p>Clause 4.10</p> <p>Clause 4.11</p> <p>Clause 4.12</p> <p>Clause 4.13</p> <p>Clause 4.14</p> <p>Clause 4.15</p> <p>Clause 4.16</p> <p>Clause 6</p>	A
<p>Requirements and test methods for Pressure Gauges and Pressure Switches</p>	<p>Pressure gauges</p> <p>Pressure switches</p> <p>Documentation</p> <p>Marking</p>	<p>BS EN 12094-10:2003</p> <p>Clause 4.1</p> <p>Clause 4.2</p> <p>Clause 4.3</p> <p>Clause 6</p>	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE FIGHTING SYSTEMS (GAS COMPONENTS) (cont'd)			
Requirements and test methods for mechanical weighing devices	Compliance Design Function at normal temperature Function at high and low temperature Overload Corrosion Documentation Marking	BS EN 12094-11:2003 Clause 3.1 Clause 3.2 Clause 3.3 Clause 3.4 Clause 3.5 Clause 3.6 Clause 3.7 Clause 5	A
Requirements and test methods for pneumatic alarm devices	Compliance Design Function at normal temperature Function at high and low temperature Functional reliability Operational reliability Function in cold and moist atmosphere Corrosion Stress corrosion Vibration Impact Documentation Marking	BS EN 12094-12:2003 Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.5 Clause 4.6 Clause 4.7 Clause 4.8 Clause 4.9 Clause 4.10 Clause 4.11 Clause 4.12 Clause 6	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
<p>FIRE FIGHTING SYSTEMS (GAS COMPONENTS) (cont'd)</p> <p>Requirements and Test Methods for Check Valves and Non-Return Valves</p>	<p>General design Connection threads Internal pressure Strength Flow way Leakage Impact Function and ambient temperature Corrosion Stress corrosion Vibration resistance Documentation Marking</p>	<p>BS EN 12094-13:2001</p> <p>Clause 4.1 Clause 4.2 Clause 4.3 Clause 4.4 Clause 4.5 Clause 4.6 Clause 4.7 Clause 4.8</p> <p>Clause 4.9 Clause 4.10 Clause 4.11 Clause 4.12 Clause 6</p>	<p>A & B</p>
<p>Gaseous Fire Extinguishing Systems</p>	<p>physical properties and system design</p>	<p>BS ISO 14520-1:2006</p> <p>BS ISO 14520-1:2006 (General requirements)</p> <p>BS ISO 14520-2:2006 (CF 31 extinguishment)</p>	<p>A</p>



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Gaseous Fire Extinguishing Systems (Cont)	physical properties and system design	BS ISO 14520-3:2000 (FC-2-1-8 extinguishment) BS ISO 14520-4:2000 (FC-3-1-10 extinguishment) BS ISO 14520-5:2006 (FK-5-1-12 extinguishment) BS ISO 14520-6:2006 (HCFC Blend A extinguishment) BS ISO 14520-7:2000 (HCFC Blend 124 extinguishment) BS ISO 14520-8:2006 (HFC 125 extinguishment) BS ISO 14520-9:2006 (HFC 227ea extinguishment) BS ISO 14520-10:2005 (HFC 23 extinguishment) BS ISO 14520-11:2005 (HFC 236fa extinguishment) BS ISO 14520-12:2005 (IG-01 extinguishment) BS ISO 14520-13:2005 (IG-100 extinguishment) BS ISO 14520-14:2005 (IG-55 extinguishment) BS ISO 14520-15:2005 (IG-541 extinguishment)	A
	Requirements for fire testing of fixed gaseous fire extinguishing systems	LPS 1230	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE FIGHTING SYSTEMS (SPRINKLERS)	<u>Construction and Performance Tests</u>	BS EN 12259-1:1999 Clause 4.1 Clause 4.2 Clauses 4.3 & 4.4 Clause 4.5 Clause 4.6 Clause 4.7 Clause 4.8 Clause 4.9 Clause 4.10 Clause 4.11 Clause 4.12 Clause 4.13 Clause 4.14 Clause 4.15 Clause 4.16 Clause 4.17 Clause 4.18 Clause 4.19	A
	Automatic sprinklers	Documented In-house Method LPS 1039:Issue 5.1:2005 excluding Clause 7.20	A
FIRE FIGHTING SYSTEMS (PORTABLE FIRE EXTINGUISHER)	<u>Performance Tests</u>		
	Fire extinguishing performance rating	BS 7937:2000	A
	Portable fire extinguisher tests and fire extinguishing performance rating	BS EN 3:1996, Parts 1, 2 (excluding Clause 4), 3, 4, 5 and 6 (Clause 3) BS EN 3-7:2004 (excluding Clause 9)	A
	Portable fire extinguisher tests and fire extinguishing performance rating	BS EN 3-1:1996, -1:1996 BS EN 3-7:2004 (Clauses 7 & 15)	C
	Disposable fire extinguishers (aerosol)	BS 6165:2002	



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FIRE FIGHTING SYSTEMS (VALVES, ACTUATORS, COUPLINGS, PRESSURE VESSELS, ETC)	Hydrostatic pressure to 700 bar	LPS 1024 LPS 1040 LPS 1041 LPS 1185 LPS 1186 BS 5155:1984 BS 5160:1989 BS 6755:Part 1:1986	A
FIRE PERFORMANCE EXTERNAL ROOF TESTS	Fire exposure roof test	Documented In-house Method SB8/52/T:2004, based on BS 476:Part 3:1958 FTSG Resolution 61 BS 476:Part 3:2004 CR 1187:2002 (inc Amendment 1:2004 ENV 1187:2002 DD ENV 1187:2002 (inc pr A1:December 2004	A
	Classification document of fire exposure roof tests	BS EN 13501-5: 2005	A
FLOORS	<u>Acoustical Tests</u>		
	Laboratory measurement of airborne sound insulation	BS EN 140-3:1995 BS 2750:Part 3:1995	A
	Laboratory measurement of impact sound insulation	BS EN ISO 140-6:1998 EN ISO 140-6:1998	A
	Laboratory measurements of the reduction of transmitted impact noise by floor coverings on a heavyweight standard floor	BS EN ISO 140-8:1998 EN ISO 140-8:1997	A
	Field measurement of impact sound insulation	BS EN ISO 140-7:1998 ISO 140-7:1998	B
	Field measurement of airborne sound insulation	BS EN ISO 140-4:1998 ISO 140-4:1998	B



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
FLOORS and FLAT ROOFS	Fire resistance	BS 476:Part 8:1972 Section 3 BS 476:Part 21:1987 Section 7 FTSG Resolutions 47, 67 & 70 BS EN 1365-2:2000 ISO 834:1975	A
FLOORS with RAISED ACCESS and HOLLOW CORES	Fire resistance	BS EN 1366-6: 2004	A
FURNITURE	<u>Fire Tests</u> Calorimeter measurements	Documented In-House Method No SP8/12/T:1999 and SP8/13T:1999	A
GENERAL MATERIALS	Toxicity	NES 713, Issue 3:March 1985	A
GENERAL MATERIALS FOR CONSTRUCTION OF RAILWAY PASSENGER STOCK AND BUILDING CONSTRUCTION	Smoke emission and Toxicity	NFX 70-100-1:2006 NFX 70-100-2:2006 BS 6853:1999:Annex B (All parts)* BS 6853:1999:Annex D (All parts)* *including IC procedure for Halides	A
GLAZING	Fire resistance	BS 476:Part 8:1972 Section 8 BS 476:Part 22:1987 Section 10 FTSG Resolutions 70 & 79 BS EN 1364-1:1999 BS EN 1364-1:2008 BS EN 1365-2:2000 ISO 3009:1981 ISO 3009/FDIS 3009: 2003	A
INDUSTRIAL, COMMERCIAL AND GARAGE DOORS AND GATES	Force for manual operation	BS EN 13241-1:2003 Clause 4.2.2 BS EN 14253:2000 BS EN 12604:2000 BS EN 12605:2000	A
	Mechanical resistance	BS EN 13241-1:2003 Clause 4.2.3 BS EN 12604:2000 BS EN 12605:2000	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
INDUSTRIAL, COMMERCIAL AND GARAGE DOORS AND GATES (cont'd)	Mechanical durability	BS EN 13241-1:2003 Clause 4.2.4 BS EN 12604:2000 BS EN 12605:2000	A
	Safe opening of vertically moving doors - individual test	BS EN 13241-1:2003 Clause 4.2.8 BS EN 12604:2000 BS EN 12605:2000	A
	Operating forces of power operated doors	BS EN 13241-1:2003 Clause 4.3.3 BS EN 12453:2000 BS EN 12445:2000	A
	Water tightness	BS EN 13241-1:2003 Clause 4.4.2 BS EN 12489:2000 BS EN 12425:2000	A
	Resistance to wind load - complete doors - individual or collective elements	BS EN 13241-1:2003 Clause 4.4.3 BS EN 12424:2000 BS EN 12604:2000 BS EN 12444:2000	A
	Noise	BS EN 13241-1:2003 Clause 4.4.4 BS EN ISO 140-3:1995 BS EN ISO 717-1:1997	A
	Air permeability	BS EN 13241-1:2003 Clause 4.4.6 BS EN 12427:2000 BS EN 12426:2000	A
	Durability of the performance characteristics	BS EN 13241-1:2003 Clause 4.4.7 BS EN 12605:2000	A
	Safe opening of vertically moving doors - individual test	BS EN 13241-1:2003 Clause 4.2.8 BS EN 12604:2000 BS EN 12605:2000	B
	Operating forces of power operated doors	BS EN 13241-1:2003 Clause 4.3.3 BS EN 12453:2000 BS EN 12445:2000	B



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
INTERIOR MATERIALS AIRCRAFT/VEHICLE	Flame resistance	CAA-8:Issue 2:1974 CS-25 Appendix F Part 1 Section 6 subsection 4	A
	Flammability	FMVS 302:1975 BS Au 169:1992 ISO 3795:1989 CS-25 Appendix F Part 1 Section 6 subsection 5	A
MACHINES	<u>Acoustics</u> Determination of sound power levels of noise sources using sound intensity. Measurement at discrete points	BS EN ISO 9614-1:1995	A
MARINE and OFFSHORE STRUCTURES	Fire resistance	Hydrocarbon DoE FRS 14-84:1990 IMO Resolution A754(18)1993 FTP Code [RES MSC 61(67)]	A
	FTP Code		
	Non-combustibility	IMO A472(XII):1984/93 IMO A799(19):1995	A
	Smoke and Toxicity	ISO 5659:Part 2:1994/1998/2006 IMO MSC 41 (64)	A
	Surface spread of flame	ISO 5658-2:1996 IMO ResA653(16) SOLAS/IMO 1986	A A
MATERIALS - for acoustical applications	<u>Physical Tests</u> Airflow resistance	BS EN 29053:1993	A
MODULAR SITE HUTS	Reaction to fire / fire tests	LPS 1195-1 issue 4:2005	C



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
NATURAL STONE	<u>Geological Tests</u>		
	Petrographic description	BS EN 12407:2007	A
	<u>Mechanical Tests</u>		
	Compressive strength (loads from 100 to 2000 kN)	ASTM C170-06	A
	Flexural strength (loads from 0.04 to 100kN)	ASTM C880-06	A
	Modulus of rupture (loads from 0.04 to 100kN)	ASTM C99-06	A
	Flexural strength under concentrated load (loads from 0.04 to 100kN)	BS EN 12372:2006	A
	Compressive strength (loads from 100 to 2000 kN)	BS EN 1926:2006	A
	<u>Physical Tests</u>		
	Bulk specific gravity	ASTM C97-02	A
	Water absorption	ASTM C97-02	A
	Water absorption	BR141:1989	A
	Saturation coefficient	BR141:1989	A
	Porosity	BR141:1989	A
	Open porosity and apparent density	BS EN 1936:2006	A
	Salt crystallisation	Documented In-House Method SP8/03/J:December 1998	A
	<u>Physical Tests</u>		
Sulphuric acid immersion	BR141:1989	A	



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
NATURAL STONE (Cont'd)	Unpolished slip resistance	BS EN 1341:2001, ANNEX D	A
	Slip resistance by means of the pendulum tester	BS EN 14231:2003	A&B
	Resistance to salt crystallisation	BS EN 12370:1999	A
	<u>Mechanical Tests</u>		
	Frost resistance	BS EN 12371:2001	A
PLASTICS, RUBBERS and RELATED MATERIALS	Rate of burn	BS 2782:Part 1:1992, Method 140A Documented In-house Method based on BS 2782:Part 5:1970, Method 508A	A
	Flammability	BS 2782:Part 1:Method 140D:1997 BS 2782:Part 1:Method 140E:1982(1988)	A
	Horizontal burn characteristics	BS 4735:1974(1984) ASTM/D 635-96	A
	Vertical burn characteristics	ASTM/D 3014-94	A
	Specific optical density of smoke	BS 6401:1983	A
PROTECTIVE COVERING MATERIALS	<u>Large Fire Tests</u>		
	Smoke/toxicity	LPS 1207:Issue 2:2003: Documented in-house method based on prEN 2824/5/6	A
RAILWAY ROLLING STOCK AND INFRASTRUCTURE	Fire tests	BS 6853:1999 (inc Amd 1) CEN TS 45545-2:2006	A
RESILIENT MATERIALS - used under floating floors	<u>Acoustical Tests</u>		
	Dynamic stiffness	BS EN 29052-1:1992 ISO 9052-1:1989	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ROOFING SLATES	<u>Chemical Tests</u> Sulphuric acid immersion	BS 680:Part 2:1971	A
	<u>Physical Tests</u> Water absorption	BS 680:Part 2:1971	A
	Wetting and drying	BS 680:Part 2:1971	A
SANDWICH PANELS	Reaction to fire / fire tests	LPS 1181-1 issue 1.1:2005 LPS 1181-2 issue 2.0:2005	A, B & C
SECURITY EQUIPMENT	Alarm systems: Intrusion & hold-up systems: Control & indicating equipment	DD CLC/TS 50131-3:2003 BS EN 50131-3:2009 excluding Water Ingress (IPX2, IPX4 & IPX7)	A
	Alarms systems: Intrusion systems: Power supplies,	BS EN 50131-6:2008 excluding Water Ingress (IPX2, IPX4 & IPX7)	A
SERVICE INSTALLATIONS Fire dampers, penetration seals, linear gap seals, closures and conveyors	Fire resistance	BS EN 1366-2: 1999 BS EN 1366-3: 2004 BS EN 1366-3: 2009 BS EN 1366-4: 2006 ISO 10294-1: 1996 BS ISO 10294-5:2005 BS EN 1366-7: 2004 AS4072-1:2005 inc Amd 1;2006 ASTM E A814:2008b	A
	Classification of fire resistance of building service installations	BS EN 13501-3: 2005+A1:2009	A
SCAFFOLD CLADDING MATERIALS	Fire tests	LPS 1215:Issue 3:2003	A
STORAGE UNITS	Fire resistance	BS EN 15659:2009	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
SUSPENDED CEILINGS PROTECTING STEEL BEAMS	Fire resistance	BS 476:Part 8:1972 Section 6 BS 476:Part 23:1987 Section 5 (including Amd 9458:1998) FTSG Resolutions 68 &70 ISO 834:1975	A
TEXTILES AND FILMS	Flammability in vertical orientation	IMO Resolution A.563(14): 1998	A
UPHOLSTERED SEATING and BEDDING - INCLUDING COMPOSITE FILLINGS FOR FURNITURE AND PILLOWS AND CUSHIONS WITH PRIMARY COVERS AND INCLUDING FOAM CUSHIONS	Ignitability	BS EN 597-1:1995 BS EN 597-2:1995 BS 5852:Part 1:1979 BS 5852:Part 2:1982 BS 5852:1990 BS EN 1021-1:1994 BS EN 1021-2:1994 IMO A688(17)	A
	Ignitability Tests	Schedule 2:Part II	A
	Ignitability Tests	Schedule 2:Part II Furniture and Furnishings (Fire) (Safety) Regs: 1988 SI. 1324	A
UPHOLSTERED SEATING and BEDDING - INCLUDING COMPOSITE FILLINGS FOR FURNITURE AND PILLOWS AND CUSHIONS WITH PRIMARY COVERS AND INCLUDING FOAM CUSHIONS Cont'd)	Ignitability Tests	Schedule 2:Part IV	A
	Cigarette Test	Schedule 4	A
POLYURETHANE FOAM SLABS and CUSHIONS	Ignitability Tests	Furniture and Furnishings (Fire) (Safety) Regs: 1988:SI.1324 Schedule 1:Part I	A



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
POLYURETHANE FOAM in CRUMB FORM	Ignitability Tests	Schedule 1:Part II	A
LATEX RUBBER FOAM	Ignitability Tests	Schedule 1:Part III	A
NON-FOAM FILLING MATERIALS, SINGLY	Ignitability Tests	Schedule 2:Part I	A
VENTILATION SERVICE DUCTS	Fire resistance	BS 476:Part 24:1987 FTSG Resolution 70 ISO 6944:1985 BS EN 1366-1: 1999 BS EN 1366-5: 2003	A
	Classification of fire resistance of building service installations	BS EN 13501-3: 2005:+A1:2009	A
VOLATILE ORGANIC COMPOUNDS - in air	<u>Chemical Analysis</u>		
	Determination of volatile organic compounds by thermal desorption, gas chromatography, mass spectrometry and FID	Documented in-house method SP8/18/W based on ISO 16000-6:2004	A
	Determination of Formaldehyde	Documented In-house Method SP8/20/W using liquid desorption and HPLC. Method based on ISO 16000-3:2001 & MDHS 78	A
	<u>Sampling</u>		
	Pumped sampling for volatile organic compounds in air	Documented in-house method SP8/19/W based on ISO 16000-6:2004	B
	Pumped sampling for Formaldehyde	Documented in house method SP/8/21/W based on ISO 16000- 3:2001	B



0578

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WALLS, PARTITIONS (smoke barriers)	Fire resistance	BS 476:Part 8:1972 Section 2 BS 476:Part 21:1987 Section 8 FTSG Resolutions 47, 60, 70 & 79 BS 476:Part 22:1987 Section 5 BS EN 1364-1:1999 BS EN 1365-1:1999 ISO 834:1975 prEN 12101-1: 2002 (section 5.2 only)	A
WALLS & PARTITIONS	<u>Acoustical Tests</u> Measurement of airborne sound insulation	BS EN ISO 140-3:1995 BS 2750:Part 3:1995	A
	Field measurement of airborne sound insulation	BS EN ISO 140-4:1998 ISO 140-4:1998	B
WALL TIES	<u>Acoustical Tests</u> Dynamic stiffness	Documented In-House Method SP8/30/V (September 2003)	A
WINDOWS – casement, top swing, vertical sliding and tilt/turn for domestic applications	<u>Physical Tests</u> Glazing removal - manual Glazing removal - mechanical Manipulation Mechanical loading Manual check	BS 7950:1997 Annex A	A



0578
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BRE GLOBAL LIMITED
Incorporating LPC Testing, BRE Testing
Issue No: 067 Issue date: 25 June 2010

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WINDOWS	Air permeability	BS 5368:Part1:1976 EN 42	A
	Watertightness under static pressure	BS 5368:Part 2:1980 EN 86	A
	Wind resistance	BS 5368:Part 3:1978 EN 77	A
	Weather tightness	BS 6375-1:2009	A
WOOD & WOOD PRESERVATIVES	<u>Biological Tests</u> Determination of protective effects using:		A
	Hylotrupes bajulus	In-house documented procedure SP6/07/K (based on BS EN 47:2005) & SP6/06/K (based on EN 46-1:2009)	
	Reticulitermes santonensis	In-house documented procedure SP6/09/K (based on BS EN 118:2005) & SP6/05/K (based on BS EN 117:2005)	A
	Basidiomycete wood destroying	In-house documented procedure SP6/04/K (based on BS EN 113:1997)	A
	Effectiveness against soft rotting micro-fungi and other soil inhabiting micro-organisms	In house documented procedure SP6/08/K (based on DD ENV 807:2001)	A
Accelerated ageing of treated wood prior to biological testing	Pre-conditioning for the above tests:- Where Evaporative ageing is required: In house documented procedure SP6/02/K (based on EN73) Where Leaching is required: In house procedure SP6/03/K (based on EN84)	A	

END